

VERMONT FORENSIC LABORATORY

103 South Main Street  
PO BOX 47  
WATERBURY, VERMONT 05676-0047  
TEL: 802-244-8788  
FAX: 802-241-5557  
<http://vfl.vermont.gov>



**Certificate of Preparation of DataMaster Interference Solution**

On 7/29/15, an interference solution was prepared by Amanda Bolduc for use as a DataMaster Interference Check solution. The solution was prepared per standard laboratory practice and assigned the lot number 15-04-08A1B and an expiration date of 1/21/16.

This solution was prepared by adding 100  $\mu$ L of acetone to 1 L of ethanol solution 15-04-080, which was certified on 1/29/15.

This solution is certified as presenting 0.080 g/210L simulated breath ethanol 0.01% vol/vol Acetone concentration and approved for use as a DataMaster Interference Check solution.

Dated at Waterbury, Vermont this 29<sup>th</sup> day of July, 2015.

  
Amanda Bolduc

Reviewed By:  Date: 7/29/2015